

Name of Accreditation Program	JCSS Accreditation Program
Accreditation No.	JCSS 0070
Date of Initial Accreditation	1996-11-14
Latest Date of Issue	2018-11-22
Name and Address of Accredited Organization	Calibration Service Center, Shinko Inspection & Service Co., Ltd. 2-3-1, Shinhama, Arai-cho, Takasago-shi, Hyogo 676-8670, Japan JCN 2140001043940
Inquiry Point	Calibration Service Center Tel: +81-79-445-7254      FAX: +81-79-445-7378
Accreditation Standards	ISO/IEC 17025:2005 (Calibration)
Accreditation Scope	As attached

\*JCN: Japan Corporate Number

General Field of Calibration: Length

Date of Initial Accreditation of the Field: 1996-11-14

Permanent Laboratory/On-site Calibration: Permanent Laboratory

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range		Expanded Uncertainty (Level of Confidence Approximately 95 %)
Length Measuring Instrument	Gauge Blocks (Comparison method)	From 0.5 mm up to 100 mm		0.09 $\mu\text{m}$
	Callipers	Up to 600 mm		0.03 mm
	Micrometers	Outside Micrometers	Up to 100 mm	3 $\mu\text{m}$
			Up to 200 mm	4 $\mu\text{m}$
			Up to 300 mm	5 $\mu\text{m}$
			Up to 400 mm	6 $\mu\text{m}$
			Up to 500 mm	7 $\mu\text{m}$

#All Calibration Procedures are in-house procedures developed by this laboratory.