

Name of Accreditation Program	JCSS Accreditation Program
Accreditation No.	JCSS0070
Date of Initial Accreditation	1996-11-14
Latest Date of Issue	2018-06-05
Name and Address of Accredited Organization	Calibration Service Center, Shinko Inspection & Service Co., Ltd. 2-3-1 Shinhama, Arai-cho, Takasago-shi, Hyogo 676-8670, Japan JCN 2140001043940
Inquiry Point	Calibration Service Center Tel: +81-79-445-7254      FAX: +81-79-445-7378
Accreditation Standards	ISO/IEC 17025:2005 (Calibration)
Accreditation Scope	As attached

\*JCN : Japan Corporate Number

General Field of Calibration : Length

Date of Initial Accreditation of the Field : 1996-11-14

Permanent Laboratory/On-site Calibration : Permanent Laboratory

Type of service		Calibration scope	CMC (Level of Confidence Approximately 95 %)
Length Measuring Instrument	Gauge Blocks (Comparison method)	From 0.5 mm up to 100 mm	0.09 $\mu\text{m}$