

Name of Accreditation Program	JCSS Accreditation Program
Accreditation Identification	JCSS 0070 Calibration
Name of Conformity Assessment Body	Calibration Service Center, Shinko Inspection & Service Co., Ltd.
Name of Legal Entity	Shinko Inspection & Service Co., Ltd. JCN 2140001043940
Inquiry Point	Calibration Service Center TEL: +81-79-445-7254 FAX: +81-79-445-7378

*JCN: Japan Corporate Number



20·05·27-NITE-011
2021-05-18

Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a calibration laboratory of Japan Calibration Service System.

Accreditation Identification: JCSS 0070 Calibration

Name of Conformity Assessment Body: Calibration Service Center,
Shinko Inspection & Service Co., Ltd.

Name of Legal Entity: Shinko Inspection & Service Co., Ltd.

Location of Conformity Assessment Body: 2-3-1, Shinhama, Arai-cho, Takasago-shi, Hyogo
676-8670, JAPAN

Scope of Accreditation: Length (as the following pages)

Accreditation Requirement: ISO/IEC 17025:2017*

* The relevant accreditation requirements described in the JCSS Accreditation Scheme Document are also applied.

Effective Date of Accreditation: 2021-05-18

Expiry Date of Accreditation: 2025-05-17

Date of Initial Accreditation: 1996-11-14

A handwritten signature in black ink, appearing to read 'Kozo Sakamoto', is written over a horizontal line.

SAKAMOTO Kozo

Chief Executive, International Accreditation Japan (IAJapan)

National Institute of Technology and Evaluation

-
- International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).
 - MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.
 - This laboratory fulfills ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation means this laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).
 - The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

General Field of Calibration: Length

Date of Initial Accreditation of the Field: 1996-11-14

Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facility

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range		Expanded Uncertainty (Level of Confidence Approximately 95 %)
Length Measuring Instrument	Gauge Blocks (Comparison method)	From 0.5 mm up to 100 mm		0.09 μm
	Callipers	Up to 600 mm		0.03 mm
	Micrometers	Outside Micrometers	Up to 100 mm	3 μm
			Up to 200 mm	4 μm
			Up to 300 mm	5 μm
			Up to 400 mm	6 μm
Up to 500 mm			7 μm	

#All Calibration Procedures are in-house procedures developed by this laboratory.