

| | |
|---|---|
| Name of Accreditation Program | JCSS Accreditation Program |
| Accreditation No. | JCSS0083 |
| Date of Initial Accreditation | 2002-06-25 |
| Latest Date of Issue | 2014-08-28 |
| Name and Address of Accredited Organization | Calibration Laboratory, Sukegawa Electric Co., Ltd. 3333-23 Kamitezuna, Takahagi-shi, Ibaraki 318-0004, Japan JCN 6050001023279 |
| Inquiry Point | Calibration Laboratory Tel: +81-293-22-0389 FAX: +81-293-22-0383 |
| Accreditation Standards | ISO/IEC 17025:2005 (Calibration) |
| Accreditation Scope | As attached |

*JCN : Japan Corporate Number

General Field of Calibration : Temperature

Date of Initial Accreditation of the Field : 2002-06-25

Permanent Laboratory/On-site Calibration : Permanent Laboratory

| Type of Service | | Calibration Scope | CMC (Level of Confidence Approximately 95 %) | | |
|---|---|-----------------------------------|--|---------------------|---------------------|
| Contact Type Thermometer | Fixed point apparatus | Triple point of water | 2.5 mK | | |
| | | Freezing point of Tin | 5.6 mK | | |
| | | Freezing point of Zinc | 7.0 mK | | |
| | Resistance thermometer (Fixed point calibration) | | | $W(T_{90})$ (*1) | $R(T_{90})$ (*2) |
| | | Triple point of water | — | 3 mK | |
| | | Freezing point of Tin | 7.8 mK | — | |
| | Resistance thermometer (Comparative calibration) | Freezing point of Zinc | 10.1 mK | — | |
| | | From -40 °C up to 35 °C | 6 mK | — | |
| | | More than 35 °C up to 200 °C | 9 mK | — | |
| | Thermocouple (Fixed point calibration) | R | From -40 °C up to 35 °C | 6 mK | — |
| | | | More than 35 °C up to 200 °C | 9 mK | — |
| | | S | More than 200 °C up to 420 °C | 13 mK | — |
| | | | Freezing point of Tin | 0.21 K | |
| | | B | Freezing point of Zinc | 0.30 K | |
| | | | Freezing point of Tin | 0.21 K | |
| | K,N,E,J,T | Freezing point of Zinc | 0.20 K | | |
| | | Freezing point of Tin | 0.48 K | | |
| | Thermocouples (Comparative calibration) | R | Freezing point of Zinc | 0.30 K | |
| | | | Freezing point of Tin | 0.24 K | |
| | | | Freezing point of Zinc | 0.46 K | |
| T | | From -40 °C up to 420 °C (*3) | 0.2 K | — | |
| | | From 200 °C up to 1100 °C (*4) | 0.1 K | — | |
| | | From 200 °C up to 1100 °C (*4) | 0.2 K | — | |
| Thermometer with indicator (Comparative calibration) | R,S,B | From -40 °C up to 200 °C | 0.018 K | | |
| | | More than 200 °C up to 420 °C | 0.020 K | | |
| | K,E,J | From -40 °C up to 420 °C(*3) | 0.3 K | | |
| N | | From 200 °C up to 1100 °C(*5) | 0.8 K | | |

(*1) Temperature converted from the ratio of the resistance $R(T_{90})$ to $R(273.16K)$, $W(T_{90})$ (*2) Temperature converted from resistance $R(T_{90})$

(*3) Calibration using working standard of platinum resistance thermometer

(*4) Calibration using standard of thermocouple

(*5) Calibration using working standard of thermocouple