

Name of Accreditation Program	JCSS Accreditation Program
Accreditation Identification	JCSS 0102 Calibration
Date of Initial Accreditation	2001-12-25
Effective Date of Accreditation	2019-12-12
Expiry Date of Accreditation	2023-12-11
Name and Location of Conformity Assessment Body	KAWAJU AKASHI ENGINEERING CO., LTD. 1-1 Kawasaki-cho, Akashi city, Hyogo-pref., 673-8666 Japan
Name of Legal Entity	KAWAJU AKASHI ENGINEERING CO., LTD. JCN 5140001034739
Inquiry Point	Calibration Section, Quality Assurance Dept. Tel: +81-78-926-1961 FAX: +81-78-926-1935
Accreditation Requirements	ISO/IEC 17025:2017 and Accreditation Requirements in the Section 6 of Accreditation Scheme (JCSS) 2nd Edition (Calibration)
Accreditation Scope	As attached

*JCN: Japan Corporate Number

General Field of Calibration: Length

Date of Initial Accreditation of the Field: 2001-12-25

Permanent Laboratory/On-site Calibration: Permanent Laboratory

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)
Length Measuring Instrument	Gauge Blocks by Comparison method	From 0.5 mm up to 100 mm	0.09 μm
	Calipers	Up to 1000 mm	0.07 mm
	Height Gauges	Up to 1000 mm	0.05 mm
	Depth Gauges	Up to 300 mm	0.03 mm
	Micrometers	Up to 500 mm	6 μm
	Indicating micrometers	Micrometer Up to 500 mm	3 μm

#All Calibration Procedures are in-house procedures developed by this laboratory.