

Name of Accreditation Program	JCSS Accreditation Program
Accreditation Identification	JCSS 0114 Calibration
Name of Conformity Assessment Body	Tokyo Metropolitan Government Inspection Institute of Weights & Measures
Name of Legal Entity	Tokyo Metropolitan Government Inspection Institute of Weights & Measures JCN 8000020130001
Inquiry Point	Verification Section Mass and Pressure Gauge Sub-section TEL: +81-3-5617-6632      FAX: +81-3-5617-6634

\*JCN: Japan Corporate Number



20·02·06-NITE-004  
2 0 2 0 - 1 0 - 1 2

## Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a calibration laboratory of Japan Calibration Service System.

Accreditation Identification: JCSS 0114 Calibration

Name of Conformity Assessment Body: Tokyo Metropolitan Government Inspection Institute of Weights & Measures

Name of Legal Entity: Same as above

Location of Conformity Assessment Body: 3-3-41, Shinsuna, Koto-ku, Tokyo 136-0075, Japan

Scope of Accreditation: Mass (as the following pages)

Accreditation Requirement: ISO/IEC 17025:2017\*

\* The relevant accreditation requirements described in the JCSS Accreditation Scheme Document are also applied.

Effective Date of Accreditation: 2020-10-12

Expiry Date of Accreditation: 2024-10-11

Date of Initial Accreditation: 2002-08-19



KISHIMOTO Isao

Chief Executive, International Accreditation Japan (IAJapan)

National Institute of Technology and Evaluation

- 
- International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).
  - MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.
  - This laboratory fulfills ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation means this laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).
  - The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

General Field of Calibration: Mass

Date of Initial Accreditation of the Field: 2002-08-19

Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facility

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range	Expanded Uncertainty (Level of Confidence Approximately 95 %) (Conventional Mass)
Weight	Weight	1000 kg	1.9 g
		500 kg	0.9 g
		200 kg	1.1 g
		100 kg	0.6 g
		50 kg	0.4 g
		20 kg	14 mg
		10 kg	6.4 mg
		5 kg	3.6 mg
		2 kg	1.5 mg
		1 kg	0.53 mg
		500 g	0.27 mg
		200 g	0.11 mg
		100 g	0.049 mg
		50 g	0.032 mg
		20 g	0.025 mg
		10 g	0.023 mg
		5 g	0.015 mg
		2 g	0.012 mg
		1 g	0.0091 mg
		500 mg	0.0058 mg
		200 mg	0.0047 mg
		100 mg	0.0036 mg
		50 mg	0.0033 mg
		20 mg	0.0027 mg
10 mg	0.0024 mg		
5 mg	0.0020 mg		
2 mg	0.0020 mg		
1 mg	0.0019 mg		

#All Calibration Procedures are in-house procedures developed by this laboratory.