

Name of Accreditation Program	JCSS Accreditation Program
Accreditation No.	JCSS0114
Date of Initial Accreditation	2002-08-19
Latest Date of Issue	2018-08-30
Name and Address of Accredited Organization	Tokyo Metropolitan Government Inspection Institute of Weights & Measures 3-3-41, Shinsuna, Koto-ku, Tokyo 136-0075, Japan JCN 8000020130001
Inquiry Point	Verification Section Mass and Pressure Gauge team Tel: +81-3-5617-6632 FAX: +81-3-5617-6634
Accreditation Standards	ISO/IEC 17025:2005 (Calibration)
Accreditation Scope	As attached

*JCN: Japan Corporate Number

General Field of Calibration: Mass

Date of Initial Accreditation of the Field: 2002-08-19

Permanent Laboratory/On-site Calibration: Permanent Laboratory

Calibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range	Expanded Uncertainty (Level of Confidence Approximately 95 %) (Conventional Mass)
Weight	Weight	1000 kg	1.9 g
		500 kg	0.9 g
		200 kg	1.1 g
		100 kg	0.6 g
		50 kg	0.4 g
		20 kg	14 mg
		10 kg	6.4 mg
		5 kg	3.6 mg
		2 kg	1.5 mg
		1 kg	0.53 mg
		500 g	0.27 mg
		200 g	0.11 mg
		100 g	0.049 mg
		50 g	0.032 mg
		20 g	0.025 mg
		10 g	0.023 mg
		5 g	0.015 mg
		2 g	0.012 mg
		1 g	0.0091 mg
		500 mg	0.0058 mg
		200 mg	0.0047 mg
		100 mg	0.0036 mg
		50 mg	0.0033 mg
20 mg	0.0027 mg		
10 mg	0.0024 mg		
5 mg	0.0020 mg		
2 mg	0.0020 mg		
1 mg	0.0019 mg		

#All Calibration Procedures are in-house procedures developed by this laboratory.