

| | |
|---|---|
| Name of Accreditation Program | JCSS Accreditation Program |
| Accreditation No. | JCSS0114 |
| Date of Initial Accreditation | 2002-08-19 |
| Latest Date of Issue | 2018-08-30 |
| Name and Address of Accredited Organization | Tokyo Metropolitan Government Inspection Institute of Weights & Measures 3-3-41, Shinsuna, Koto-ku, Tokyo 136-0075, Japan JCN 8000020130001 |
| Inquiry Point | Verification Section Mass and Pressure Gauge team Tel: +81-3-5617-6632 FAX: +81-3-5617-6634 |
| Accreditation Standards | ISO/IEC 17025:2005 (Calibration) |
| Accreditation Scope | As attached |

*JCN : Japan Corporate Number

General Field of Calibration : Mass

Date of Initial Accreditation of the Field : 2002-08-19

Permanent Laboratory/On-site Calibration : Permanent Laboratory

| Calibration Procedures# and Type of Instruments/Materials to be calibrated | | Range | CMC (Level of Confidence Approximately 95 %) (Conventional Mass) |
|--|-----------|---------|---|
| Weight | Weight | 1000 kg | 1.9 g |
| | | 500 kg | 0.9 g |
| | | 200 kg | 1.1 g |
| | | 100 kg | 0.6 g |
| | | 50 kg | 0.4 g |
| | | 20 kg | 14 mg |
| | | 10 kg | 6.4 mg |
| | | 5 kg | 3.6 mg |
| | | 2 kg | 1.5 mg |
| | | 1 kg | 0.53 mg |
| | | 500 g | 0.27 mg |
| | | 200 g | 0.11 mg |
| | | 100 g | 0.049 mg |
| | | 50 g | 0.032 mg |
| | | 20 g | 0.025 mg |
| | | 10 g | 0.023 mg |
| | | 5 g | 0.015 mg |
| | | 2 g | 0.012 mg |
| | | 1 g | 0.0091 mg |
| | | 500 mg | 0.0058 mg |
| | | 200 mg | 0.0047 mg |
| | | 100 mg | 0.0036 mg |
| | | 50 mg | 0.0033 mg |
| 20 mg | 0.0027 mg | | |
| 10 mg | 0.0024 mg | | |
| 5 mg | 0.0020 mg | | |
| 2 mg | 0.0020 mg | | |
| 1 mg | 0.0019 mg | | |

#All Calibration Procedures are in-house procedures developed by this laboratory.