

Name of Accreditation Program	JCSS Accreditation Program
Accreditation No.	JCSS0150
Date of Initial Accreditation	2004-10-20
Latest Date of Issue	2015-06-19
Name and Address of Accredited Organization	Balance Calibration Division, SHIMADZU ACCESS Corporation 20-8 Asakusabashi, 5-Chome, Taitou-ku, Tokyo 111-0053, Japan JCN 7010501009474
Inquiry Point	Balance Calibration Division Tel: +81-03-5820-3277 FAX: +81-03-5820-3275
Accreditation Standards	ISO/IEC 17025:2005 (Calibration)
Accreditation Scope	As attached

*JCN : Japan Corporate Number

General Field of Calibration : Mass

Date of Initial Accreditation of the Field : 2004-10-20

Permanent Laboratory/On-site Calibration : Permanent Laboratory, On-site Calibration

Type of Service		Calibration Scope	CMC (Level of Confidence Approximately 95 %)	
			Permanent Laboratory	On-site Calibration
Scale	Non-automatic Electronic Weighing Instruments	From 1 g up to 5 g	0.07 mg	0.07 mg
		More than 5 g up to 10 g	0.13 mg	0.13 mg
		More than 10 g up to 50 g	0.29 mg	0.30 mg
		More than 50 g up to 100 g	0.39 mg	0.41 mg
		More than 100 g up to 150 g	0.34 mg	0.37 mg
		More than 150 g up to 220 g	0.44 mg	0.50 mg
		More than 220 g up to 300 g	0.62 mg	0.69 mg
		More than 300 g up to 400 g	0.78 mg	0.88 mg
		More than 400 g up to 490 g	0.93 mg	1.1 mg
		More than 490 g up to 1000 g	3.6 mg	3.7 mg
		More than 1 kg up to 2 kg	20 mg	20 mg
		More than 2 kg up to 5 kg	24 mg	25 mg
		More than 5 kg up to 10 kg	36 mg	37 mg
		More than 10 kg up to 20 kg	0.2 g	0.2 g
		More than 20 kg up to 34 kg	0.21 g	0.21 g
		More than 34 kg up to 50 kg	1.6 g	1.6 g
		More than 50 kg up to 100 kg	3.6 g	3.6 g
More than 100 kg up to 150 kg	4.9 g	4.9 g		