

Name of Accreditation Program	JCSS Accreditation Program
Accreditation No.	JCSS0165
Date of Initial Accreditation	2005-09-01
Latest Date of Issue	2015-03-02
Name and Address of Accredited Organization	Taisho Balance Mfg. Co., Ltd. 4219-72, Takasai, Shimotsuma-shi, Ibaraki 304-0031, Japan JCN 9050001013351
Inquiry Point	Calibration Department Tel: +81-296-43-7021    FAX: +81-296-43-8150
Accreditation Standards	ISO/IEC 17025:2005 (Calibration)
Accreditation Scope	As attached

\*JCN : Japan Corporate Number

General Field of Calibration : Mass

Date of Initial Accreditation of the Field : 2005-09-01

Permanent Laboratory/On-site Calibration : Permanent Laboratory

Type of Service		Calibration Scope	CMC(Level of Confidence Approximately 95 %)	
			Conventional Mass	
Weight	Weight	20 kg	22 mg	
		10 kg	9.4 mg	
		5 kg	3.8 mg	
		2 kg	1.6 mg	
		1 kg	0.64 mg	
		500 g	0.32 mg	
		200 g	0.16 mg	
		100 g	0.072 mg	
		50 g	0.050 mg	
		20 g	0.042 mg	
		10 g	0.030 mg	
		5 g	0.024 mg	
		2 g	0.016 mg	
		1 g	0.013 mg	
		500 mg	0.0094 mg	
		200 mg	0.0078 mg	
		100 mg	0.0064 mg	
		50 mg	0.0048 mg	
		20 mg	0.0040 mg	
		10 mg	0.0036 mg	
		5 mg	0.0036 mg	
		2 mg	0.0036 mg	
		1 mg	0.0036 mg	
		Deadweight	More than 20 kg up to 30 kg	32 mg
			More than 10 kg up to 20 kg	22 mg
			More than 5 kg up to 10 kg	10 mg
More than 2 kg up to 5 kg	5.2 mg			
More than 1 kg up to 2 kg	1.6 mg			
More than 200 g up to 1 kg	0.84 mg			
From 10 g up to 200 g	0.30 mg			